

RIKEN	PORTABLE HC/O ₂ MONITOR
-------	------------------------------------

MODEL R X - 4 1 5

OPERATION MANUAL

MARINE USE

For HC gas use	Model RX-415 (TYPE HC) Measuring range : HC : 0-100%LEL, 0-100vol% O ₂ : 0-25vol%
For CH ₄ gas use	Model RX-415 (TYPE CH ₄) Measuring range : CH ₄ : 0-100%LEL, 0-100vol% O ₂ : 0-25vol%

Precaution in operation

- After reading this manual well, start to operate this instrument.
- Keep this manual where to carry out easily
- When take out this manual by use of construction work, be sure to put it where it was.
- Do not use this instrument for no other than given purpose.
- When operate this instrument without using this manual or repair it by use of no other parts than the genuine one, the safety and quality of products could not be guaranteed. When any accident should take place by those irregular method, we cannot assume the responsibility for it.



RIKEN KEIKI CO., LTD.

2-7-6 Azusawa Itabashi-ku Tokyo, 174-8744 Japan

Phone : Tokyo (03) 3966-1113
 Telex : 272 2638 RKNFNE
 Fax : (03) 3558 9110 G
 E-Mail : intodept@rikenkeiki.co.jp

In the beginning

It is of our great pleasure to purchase Riken portable HC/O₂ gas detector model RX-415 this time. This instrument is an explosion-proof type portable gas detector which is designed to measure the presence of crude vapors in an inert gas or in air such as HC gas, vapors and methane gas and O₂.

This instruction manual is a guide book for operation of Riken portable HC/O₂ gas detector model RX-415. It is kindly requested to read and understand this content by experienced users as well as beginners. The following headline shall be shown to carry out the safety and effective work in this instruction manual.

Danger

This means that it gives the serious harm to the human life, body or material directly such as touching high voltage.

Warning

This means that it gives the serious harm to the human body or material if do not perform the operation according to the instruction manual.

Caution

This means that it gives the slight harm to the human body or material if do not perform the operation according to the instruction manual.

*** Note**

This means the advice in the operation.



Model RX-415 consists of two type according to the kind of gases below ;

For HC gas use :	Model RX-415 (TYPE HC)
	HC : 0-100%LEL/0-100vol%
	O ₂ : 0-25vol%
For CH ₄ gas use :	Model RX-415 (TYPE CH ₄)
	CH ₄ : 0-100%LEL/0-100vol%
	O ₂ : 0-25vol%

Contents

	Page
1 . Name of each part	3
2 . Operation	
2 - 1 . Preparation	5
2 - 2 . Start-up procedure	7
2 - 3 . Check of flow monitors	8
2 - 4 . Zero adjustment	8
2 - 5 . Measurement	9
2 - 6 . Stop of pump operation	11
2 - 7 . Measurement completion	12
3 . Alarm function	
3 - 1 . Kind of alarm and alarm pattern	13
3 - 2 . Countermeasure at alarm	14
4 . Maintenance check	
4 - 1 . Battery replacement	15
4 - 2 . Gas calibration	16
4 - 3 . Filter replacement	21
4 - 4 . O2 sensor replacement	22
4 - 5 . Daily check and regular maintenance check	23
4 - 6 . Replacement of parts	23
4 - 7 . Storage and treatment when not to use for a long time	23
5 . Scrap of instrument	24
6 . Trouble-shooting	25
7 . Caution in operation	26
8 . Definition of words	27
9 . Specifications	
9 - 1 . Specifications	28
9 - 2 . Accessories	30
1 0 . Detection principle	31

1. Name of each part



Power/Enter switch

This is used for the confirmation of power ON, OFF/input.

PUMP/(+) switch

This is used for pump Switch ON/OFF and to increase the input value.

ZERO/(-) switch

This is used for zero adjustment and decrease the input value.

PEAK switch

This is used when desired to show the peak value.

BATT/ESC switch

This is used when desired to show the battery voltage and cancel the input.

Gas outlet

This is the outlet to exhaust sample gas out.

Calibration gas name plate

This shows the calibration gas name.

Display unit (LCD display with back light)

This displays gas concentration.

Alarm lamp

This flickers and lights on at trouble alarm time.

Photo sensor window

This is for automatic back-lights of display unit by detecting that it gets dark around.

Gas inlet

This is a measuring gas inlet.

Carrying case

Junction tube (20cm)

Gas sampling relay hose (29m) + Gas sampling hose (1m)

Filter tube with flowmonitor

2. Operation

Check item before operation

Instrument :

- Check that there is no damage on the display unit.

Filter tube with flow monitor :

- Check the dirt of cotton and when it gets dirty, replace the cotton with new one. (See 5-3)
- When the water drops are deposited, drain it out.
- Check that there is no crack or damage.
- Check that there is no looseness on the nipple mounting.

Gas sampling hose :

- Check that there is no break or twist of it.

2 - 1 . Preparation

- (1) Mount the battery into instrument. (See 5-1)
- (2) Put the instrument into the carrying case.
- (3) Connect the instrument (gas inlet) the junction tube, filter tube with flowmonitor and gas sampling hose with instrument in turn.



As there are two types of head hoses for this instrument, use it according to the each environment.

Standard specification Gas sampling relay hose (30m) + Sampling hose Bar type (1m)


Optional specification Gas sampling relay hose (30m) + Sampling hose Floating type (1m)

(Gas sampling relay hose and sampling hose (Bar type/Float type) is optional accessories)


- Connection between gas sampling relay hose and model RX-415.

By holding knurling part with fingers, pull it to the arrow mark.

Plug in to the arrow mark with its condition. Release the fingers holding the knurling part and press in. Then, locked after ticking sound.



- Connection between gas sampling relay hose and sampling hose 3.
By holding the socket of gas sampling relay hose and the plug of sampling hose by fingers and plug in to the arrow mark until it gets a ticking sound.



⚠ Warning

- As this is used in dangerous zone, be sure to use it with carrying case.


⚠ Caution

- Be sure to plug in the gas sampling hose and junction tube.
If not surely connected, the accurate measurement cannot be carried out.
- Be sure to mount the filter tube with flowmonitor.
If measure without filter, it will be a cause of trouble by sucking dust, water and oil etc.

2 - 2 . Start-up procedure

Hold pressing power switch for 5 seconds and the power will turn on. Via battery check, filter probe connection check and sensor connection check, the gas detection will start. The sequence performance until the gas detection starts is as follows :

Hold pressing power switch for 5 seconds



* Note

At power on, check that the buzzer and pump sound.

Display, performance at pump connection failure. (See 3.)

FAIL PUMP (Lamp : Light-on, Buzzer : Continuous)

Display, performance at power voltage shortage. (See 3. and 5-1)

REPLACE BATTERY (Lamp : Light-on, Buzzer : Continuous)

Display, performance at sensor connection failure. (See 3.)

In case of HC sensor connection failure

FAIL SENSOR HC (Lamp : Light-on, Buzzer : Continuous)

In case of O₂ sensor connection failure


FAIL SENSOR O₂ (Lamp : Light-on, Buzzer : Continuous)

In case of HC sensor/O₂ sensor connection failure


FAIL SENSOR (Lamp : Light-on, Buzzer : Continuous)

2 - 3 . Check of flowmonitor

When turn on power, the pump starts working. Check that it works properly by the flowmonitor (float) of filter tube with flowmonitor.



When sucking properly (Can see the float)



Not sucking properly (Can see the float)

2 - 4 . Zero adjustment

Make zero adjustment by removing the gas sampling hose. After the detection starts, see the flow monitor and check that the air is sucked properly. Then, check that the reading gets stable and make zero adjustment by sucking the fresh air into instrument.

The operation procedure is shown below.

0 . 0 % L E L 2 0 . 9 %

(Gas display)

H O L D Z E R O K E Y

Hold pressing **Z E R O** for approx. 3 sec.

A P P L Y F R E S H A I R

After checking that fresh air is sucked into instrument and then, press **E N T E R** switch.

A D J U S T I N G Z E R O

Zero calibration failure

Approx. 5 sec.

Z E R O S E T O K

Approx. 5 sec.

0 . 0 % L E L 2 0 . 9 %

(Gas display)

Display, performance at zero calibration failure (See 3.)

In case of HC zero cal. failure

FAIL ZERO HC /

In case of O2 zero cal. failure

FAIL ZERO /O2

In case of HC/O2 zero cal. failure

FAIL ZERO HC/O2

Caution

When make zero adjustment, do it after having fresh air sucked into instrument for over 1 minute.

If zero point is adjusted under presence of gas, the accurate detection can not be carried out, and oxygen deficiency accident may be occurred.

*** Note**

- When desired to cancel the pre-operation, press **ESC** switch but after confirmed, it can not be cancelled.

2 - 5 . Measurement

(1) Measurement in tanks (Accessory 30m hose used)

Fall down to the sampling point the gas sampling hose slowly by taking care of bending and twisting of hose.

This sampling hose shows the length at every 5m and then, stop by seeing the distance at the position where desired to measure.

By the seeing the flowmonitor, check that it is sucked properly. After stopping the sampling hose at the point where desired to measure and over 3 minutes have passed, read out the display reading.

Whenever change the place of measurement, do the same procedure above.

Danger

- There may be the oxygen deficiency at gas outlet point due to the inert gas.
Do not breathe at all.
- There may be the exhaust of high density gas (Above LEL).
Do not approach the fire to it.

⚠ Warning

- When the temperature is drastically changed at next detection point, in order to make accurate measurement, wait until the instrument accustom itself to its temperature (approx 10 min.). Then, start measurement after air calibration.

⚠ Warning

- When measure O₂ in inert gas, confirm that CO₂ density of sucked air is under 15%. If over 15% of CO₂ is introduced, that may decrease the sensor life. (See 8. Caution in operation)

⚠ Caution

- Do not let water or oil sucked into instrument. If it should be sucked, the pump or sensor shall be failed.
- When measure, check the running sound of pump and flow monitors. It cannot be measured at stop of pump operation.
- When the measuring gas is except calibration gas, the display reading will show at slightly high or slightly low side position.
- Do not block the gas outlet because the display reading may show the slightly high position.

* Note

- When suck high density gas above LEL, the hang-up phenomenon takes place from the adsorption of it to the gas sampling hose and filter tube.


When use the range of vol%, use it as it is.

When make zero adjustment, remove the gas sampling hose and let fresh air sucked into instrument.

When measure by %LEL range, clean the gas sampling hose by air and try to measure after the display reading goes down zero.

(2) Battery voltage


By pressing BATT switch while in operation, the battery voltage at present can be confirmed. The battery voltage will return to the gas concentration reading if there is no input of this switch "ON" for 20 seconds.



(3) Read-out of peak value

When press PEAK switch in operation, this shows the peak value (Max point of reading for HC, Minimum point of reading for O₂) from the power switch "ON" up to now.

Then press PEAK switch again, the peak model is cancelled and returns to the measurement mode. Then if there is no input of peak for about 20 seconds, this shall return to the measurement mode. When hold pressing ESC switch for 3 seconds in the peak display, the peak value will be cleared.



2 - 6 . Stop of pump operation

When transfer the sampling point or stop measurement temporarily, it can stop the running of pump only.

When stop pump running, the battery consumption can be cut by about 40% as compared with the pump running time. All except pump work and then, no warming-up time is required if re-operate the pump.

* Note

- When do not measure for some time, turn off the power.
As all except pump work, the battery will be consumed if leave the pump operation stopped.

(1) Pump running display

< Pump running on >

This displays gas reading while in pump running.

0 . 0% L E L 2 0 . 9% (Pump running)

< Pump power off >

The following message shows while in pump work stopping.

PUMP POWER OFF (Pump stop operation)

(2) Operating procedure

Pump : Running on stop

0 . 0% L E L 2 0 . 9% (Pump running)

HOLD PUMP KEY	Press PUMP for about 3 sec.
PUMP POWER OFF	(Pump stop operation) Release PUMP switch after pump off.
Pump : Stop Operation	
PUMP POWER OFF	Pump power off
HOLD PUMP KEY	Press PUMP for about 3 sec.
PUMP POWER ON	Display for about 5 sec. Release PUMP switch after pump running.
0 . 0% L E L 2 0 . 9%	(Pump running)

2 - 7 . Measurement completion

(1) Treatment after measurement

Roll up the sampling hose in a round bundle so that any bent or twist can not be made carefully. Make air cleaning for minimum 5 seconds until the reading comes nearby to zero point under condition that the sampling hose is connected with instrument by sucking fresh air.

(2) Power OFF

While pressing the power switch for about 5 seconds, the power will be off.

While pressing the power switch off, the buzzer will sound for about 15 times.

Caution

- Do not forget to do air cleaning.

When neglect the air cleaning, the following will take place and will be a cause for trouble.

If water, oil and mist are left and condensed in dust, much of water and oil shall be sucked into instrument at the next measurement time.

As HC vapors may be absorbed completely, there will be a trouble for next time measurement.

3. Alarm function

3 - 1 . Kind of alarm and alarm function

Trouble alarm is provided for this instrument.

Each alarm is provided with buzzer, lamp and display.

Trouble alarm and alarm pattern

		Lamp	Buzzer	
Power ON	System error	ON	Continuous	" SYSTEM ERROR "
	Battery power shortage	ON	Continuous	"REPLACE BATTERY"
	Sensor disconnection	ON	Continuous	" FAIL SENSOR " " FAIL SENSOR HC " " FAIL SENSOR O2 "
	Pump disconnection	ON	Continuous	" FAIL PUMP "
Measurement	Low battery warning	Flicker	Intermittent	"B" flicker at left end
	Low battery alarm	ON	Continuous	"REPLACE BATTERY"
	ZERO cal. failure	OFF	No sound	" FAIL ZERO HC/ " " FAIL ZERO /O2" " FAIL ZERO HC/O2"
	SPAN failure	OFF	No sound	" FAIL SPAN HC/ " " FAIL SPAN /O2 " " FAIL SPAN HC/O2 "

3 - 2 . Countermeasure at alarm

(1) System error ("SYSTEM ERROR")

When receive an excessive noise, this error may appear.

(2) When the low battery alarm is given, replace 4 pcs of batteries with new ones according to the procedure of "5-1. Battery replacement".

(3) Sensor connection error

When give an excessive shock or impulse to the instrument such as dropping or throwing etc. and use for a long period, such display may appear.

Carry out the daily check or regular check (See 5. Maintenance check)

(4) Pump connection error ("FAIL PUMP")

When give an excessive shock or impulse to the instrument such as dropping or throwing etc. and use for a long period, such display may appear.

Carry out the daily check or regular check (See 5. Maintenance check)

(5) Zero cal. error ("FAIL ZERO HC/O₂")

Under the condition that zero gas(air) is not sucked correctly or water and oil etc are sucked into instrument, such display may appear. Check whether there is any damage or any sign of water or oil etc suction in the sampling hose or filter tube with flowmonitor etc. If any error could not be found, let zero gas(air) sucked into instrument correctly and make zero adjustment.

(See para 2-4)

(6) SPAN failure ("FAIL SPAN HC/O₂")

Under the condition that zero gas(air) is not sucked correctly or water and oil etc are sucked into instrument, such display may appear. Check whether there is any damage or any sign of water or oil etc suction in the sampling hose or filter tube with flowmonitor etc. If any error could not be found, let calibration gas sucked into instrument correctly and make span adjustment

(See para 2-4)

4. Maintenance check

4 - 1 . Battery replacement

Warning

- Battery replacement shall be done at non-hazardous zone by all means.
- Be sure to use the genuine batteries.

When make battery replacement, replace all 4 pcs batteries with new ones.

- (1) Check that the power OFF.

When the power is ON, do it after power OFF.



- (2) Remove the carrying case from the instrument.
- (3) Open the battery cover by turning the screw on the bottom of instrument counterclockwise with minus screwdriver or coin etc.
- (4) Remove 4 pcs batteries and put the new ones where they were by taking care of polarity.

*** NOTE**

When remove battery, remove it from the polarity of (+). When put it, it is easy to replace the battery from the polarity of (-).

- (5) When finish the replacement of batteries, put the battery cover in the reverse way.


Warning

- Tight up the battery cover for sure.

4 - 2 . Gas calibration

The measuring range of model RX-415 are dual expanded scale of 0-100%LEL and 0-100vol% for HC, and 0-25vol% for O2. Then, the gas calibration for all ranges is required.

- (1) Prepare the calibration gas (both for low density gas and high density gas) and sampling bag.
- (2) Remove the filter tube and gas sampling hoses from the instrument. Then, arrange to connect sampling bag from the gas inlet of instrument.
- (3) Turn on the power and make zero adjustment.
- (4) Put the calibration gas*¹ sampling bag and make gas calibration in the following procedure.



0 . 0% L E L 2 0 . 9% (Gas density display)

Display and function in gas calibration error. (See 3)

FAIL SPAN HC/


Danger

- Do not approach the fire at the time of gas calibration at all.
For gas calibration, use the high density gas.

*** Note**

- Before gas calibration, confirm that the indication value is “0.0%LEL” after fresh air is sucked. Be sure to make zero adjustment when the value is not “0.0%LEL”.
End even though “0.0%LEL” is displayed, if it is flickered, zero point is deviated.
Make zero adjustment.


(5) Put the calibration gas*2 to sampling bag and make gas calibration in the following procedure.



Display and function in gas calibration error.


FAIL SPAN HC/

(6) Put the calibration gas*³ to sampling bag and make gas calibration in the following procedure.



Display and function in gas calibration error.

FAIL SPAN /O2




⚠ Caution

- Do not supply the calibration gas directly by plugging in the gas inlet from the canned gas. Because the inside of instrument may be damaged.
- Be sure to make calibration by both high density gas and low density gas.
If neglect this calibration, the accurate measurement cannot be assured.

4 - 3 . Filter replacement

Check the filter (Cotton) in the filter tube with flowmonitor before measurement and if it is dirty, replace it with new one.

- (1) By turning the knurling part of the filter tube with flowmonitor, open the filter tube with flowmonitor.
- (2) Take out the cotton filter in the filter tube with flowmonitor by use of tweezers etc.
- (3) Put the new cotton filter so that it can be even in the filter tube with flowmonitor and close knurling part in the reverse procedure.



⚠ Caution

- Do not put much of cotton filter.
If put in too much, the flow rate gets down and the accurate measurement cannot be carried out.

*** Note**

- Make this filter replacement in every month.

4 - 4 . O2 sensor replacement

O2 sensor life of RX-415 is approx. 1 year. If the following phenomenon would be occurred, replace the sensor with new one. Sensor type is "OS-B3".

<Phenomenon>


When "Calibration error" is displayed, even though the air calibration is performed by fresh air.
 When the indication value can not be decreased even though human exhalation is introduced.
 When the indication is unstable. When it takes long time to get the stable indication.

Caution


- Sensor life depends on utility condition and utility frequency.
 Refrain to use the instrument under the condition that there is much dust, pressure alternation, excessive high/low temperature/humidity. These condition may decrease sensor life.

Replacement procedure

- (1) Remove the switch side cover by taking off 4 pcs of screws. (See D-1)
- (2) Pull out the connector from O2 sensor.
- (3) After removal of mounting screws and plate, O2 sensor can be taken off. (See D-2)
- (4) Take out spare O2 sensor from the case and take off copper wire from connector.
 Then, connect O2 sensor to the instrument.
- (5) Sensor shall be installed in direction as D-2 by mounting screw (2pcs) and plate.
- (6) Put the cover back and mount 4 pcs of screws. Sensor replacement is completed.



O2 sensor connector



Mounting plate and screws

4 - 5 . Daily check and regular maintenance check

(1) Daily check

- Are there any damage on switches, lamp, display and panel?
- Pump suction work check (Is pump suction sound normal? Can any trouble sound be heard?)
- Battery voltage check.

(2) Regular maintenance check

It is recommended to receive the maintenance check at least one in a year.

Contact the nearest agent or Riken Keiki Co., Ltd.

4 - 6 . Replacement of parts

As the following parts have its own end, it must be replaced regularly.

When replace, contact the nearest agent or Riken Keiki Co., Ltd.

Pump : Approx. 2 years (Depends on use frequency)

Internal filter : Approx. 1 year (Depends on use frequency)

Sensor : Approx. 5 years

4 - 7 . Storage and disposal when the instrument is not used for long time.

Storage under following conditions.

Temperature : 5 – 35

Humidity : 30 – 80RH

Place : Gas and solvent vapor are not existing.

Storage the instrument in the box where it was.

When there is no storage box, storage in vinyl and etc.

Storage the instrument indoor escaping from direct sunshine.

Take off the batteries when do not use for over 1 month.

5. Scrap of instrument

When scrap the instrument after a long use, do it in the same treatment as industrial scrap goods (Non-flammable goods) because no harmful material for environment is used except for O2 sensor. For scrap of O2 sensor, contact the nearest agent or Riken Keiki Co., Ltd.

6. Trouble shooting

This trouble shooting does not mention the possible cause of all the malfunction but simply shows it to help the cause research of probable malfunction.

Trouble	Cause	Treatment
Power can not be on	<ul style="list-style-type: none"> • Empty batteries • Batteries are consumed up. • Wrong polarity of battery • Time shortage to press Power switch 	<ul style="list-style-type: none"> • Put batteries in correct way by seeing the battery replacement. (See 5-1.) • Hold pressing for approx. 5 sec. (Until displayed)
No pump running	<ul style="list-style-type: none"> • No measurement and pump stops running (Display comment) 	<ul style="list-style-type: none"> • Make re-operation of pump (See 2-6)
No pump suction	<ul style="list-style-type: none"> • The pipe of IN, OUT of filter or tube etc may be removed or clogged. 	<ul style="list-style-type: none"> • Check the clog of filter pipe or twist etc and treat it correctly (See 5-3 and 2-5)
Zero and span adjustment for O ₂ is not available.	<ul style="list-style-type: none"> • Is sensor life period in valid? 	<ul style="list-style-type: none"> • The validity of sensor is expired. Replace it to new one.
Zero "0.0"%LEL of gas display is flickering	<ul style="list-style-type: none"> • Zero point is deviated. 	<ul style="list-style-type: none"> • Check that there is gas free around and make zero adjust. (See 2-4)
"100.0vol%" flickering	<ul style="list-style-type: none"> • Zero point or gas sensitivity is deviated. 	<ol style="list-style-type: none"> 1 . Make zero adjustment. (See 2-4) 2 . When not adjusted to reading value, make gas calibration. (See 5-2)

7. Caution in operation

Keep the following items below to maintain the function of instrument and for safety.

Warning

- Do not modify or alter the circuit or structure etc.
After modifying and altering the instrument, the function can not be maintained any longer.
- As this is explosion-proof type instrument, make the battery replacement at non-hazardous-zone. The replacement at hazardous zone will be beyond the scope of explosion-proof concept.
- As this is explosion-proof type instrument, be sure to operate with carrying case on.

Warning

- Sensor life can be excessively reduced under the condition where acid gas and solvent gas are existing. In this case, make measurement as quickly as possible.
After measurement, make sensor cleaning by introducing fresh air.

Caution

- Do not fall or crush it.
As this is a fine instrument, the function may not be maintained if high impulse or shock is given.
- Do not splash water directly.
As this is not drip-proof structure, it will be a cause for trouble if water is splashed directly.
- Do not give the electrical noise during operation by strong walkie-talkie wave etc.
When give the electrical noise to the instrument, it will be a cause to damage the instrument.
- When the temperature in air is changed drastically, it may affect the reading of instrument.
- When it is in dew or condensed, the normal measurement can not be carried out.
- For internal filter and filter tube, be sure to operate it with them on by all means.

* Note

- When measure the gas contains water, suck it after removing water.
- When measure high temperature gas, make measurement after the instrument accustom itself to its temperature.

8. Definition of words

% L E L

When define the Lower Explosion Limit of combustible gas as 100%.

Methane : 100% L E L = 5.0 v o l %

Iso-butane : 100% L E L = 1.8 v o l %

v o l %

This is the unit shown by percentage for how much a special material (or gas) in a volume is occupied in that volume.

Combustible gas

The lower limit of Explosive Limit (Explosive limit when mixed with air) is to be below 10%.
The difference between upper and lower limits is to be above 20%.

HC gas

General name of hydrocarbon.

This instrument displays it by converting through iso-butane gas density.

in Air

The atmosphere of $-10 \sim +40$, below 90%RH in an atmospheric pressure (0113hPa)

Hang-up phenomenon

This is the phenomenon to show slightly higher reading than the actual one by the influence residual gas for some while though fresh air is supplied after high density gas was sucked into instrument.

Oxygen deficiency

Phenomenon that may cause human health disorder.

9. Specifications

9 - 1 . Specifications

MODEL : R X - 4 1 5 (T Y P E H C)

Detection principle	Galvanic cell	NDIR (Non-Dispersive Infrared method)
Detection gas	Oxygen (O ₂)	HC
Calibration gas	Oxygen	Iso-buthane
Measurable range	0-25vol% (0.1vol%)	0-100%LEL / 0-100vol% [Dual auto range] (0.5%LEL) / (0.5vol%)
Increment	Within ± 0.7vol% (Follow to JIS T-8201)	± 5% of full scale (0-100%LEL) ± 5% of full scale (0-100vol%) (Under same condition)
Response time	Within 20 sec. (T ₉₀ , When sucked from gas inlet)	Within 30 sec. (T ₉₀ , When sucked from gas inlet) Within 2 min. (T ₉₀ , When sucked from end of 30m hose)
Sampling	Suction pump	
Suction rate	Over 300ml / min.	
Ambient temp/ Humidity	-10 ~ 40 / below 90%RH (Non-condensing)	
Power source	Alkaline batteries (Size C) 4 pcs	
Continuous operation time	Approx. 40 hours (No alarm / No light at 20)	
Explosion proof	Exiad BT3 (pending)	
Dimension/ weight	Approx. 200(W) x 80(H) x 142(D)mm / Approx. 2.0kg (Batt. included)	

MODEL : RX - 415 (TYPE CH₄)

Detection principle	Galvanic cell	NDIR (Non-Dispersive Infrared method)
Detection gas	Oxygen (O ₂)	CH ₄
Calibration gas	Oxygen	Methane
Measurable range	0-25vol% (0.1vol%)	0-100%LEL / 0-100vol% [Dual auto range] (0.5%LEL) / (0.5vol%)
Increment	Within ± 0.7% (Follow to JIS T-8201)	± 5% of full scale (0-100%LEL) ± 5% of full scale (0-100vol%) (Under same condition)
Response time	Within 20 sec. (T ₉₀ , When sucked from gas inlet)	Within 30 sec. (T ₉₀ , When sucked from gas inlet) Within 2 min. (T ₉₀ , When sucked from end of 30m hose)
Sampling	Suction pump	
Suction rate	Over 300ml / min.	
Ambient temp/ Humidity	-10 ~ 40 / below 90%RH (Non-condensing)	
Power source	Alkaline batteries (size C) 4 pcs	
Continuous operation time	Approx. 40 hours (No alarm / No light at 20)	
Explosion proof	Exiad BT3 (pending)	
Dimension/ weight	Approx. 200(W) x 80(H) x 142(D)mm / Approx. 2.0kg (Batt. included)	

9 - 2 . Accessories (Common for HC and CH4)

(1) Standard accessories

- Alkaline batteries (Size C) 4pcs
- Filter tube with flow monitor and connection line with coupling 1pce
- Spiral sampling tube (1m) 1pce
- Gas sampling probe 1pce
- Carrying case with shoulder strap 1pce
- Operation manual
- Test report

(2) Optional accessories

- Sampling line complete with couplings (30m) 1pce
- Absorbent cotton
- Metal storage box
- Gas sampling bag

10. Detection principle


10 - 1 . NDIR method (Non-Dispersive Infrared)

Model RX-415 is based on NDIR method (Non-Dispersive Infrared) and this structure is shown below.

The infrared beam emitted from the light source passes through the measuring cell, and optical band pass filter which can pass the absorption wave of measuring gas and attains to the infrared sensor. The amount of infrared attaining to the infrared sensor through the measuring cell and will decrease according to its density.

The variable amount of infrared is measured by the infrared sensor and it is displayed as gas concentration.

Then, there is no sensitivity against CO₂ and CO etc which have the different absorption wave from the measuring gas. Then there is no sensitivity against N₂ and H₂ etc which cannot absorb infrared. As compared with the conventional catalytic combustion method, there is no poisoning material to be absorbed and almost no sensitivity drop on this detection principle.



10 - 2 . GALVANIC CELL method

Galvanic cell is consisting of a lead anode and a gold cathode in electrolyte covered by a membrane. When oxygen enters into the gold cathode, a current which is directly proportional to the oxygen concentration will be produced, and amplified current will produce a reading on a meter in percent oxygen.

